



Sir Safety System S.p.A. Single-member company Wholly-owned subsidiary of Sir Holding Srl Via dei Fornaciai, 9 - Santa Maria degli Angeli 06081 Assisi - Perugia - Italy Tel.+39 075.8043737 - Fax +39 075.8048686 VAT Nr. IT 03359340548 Viale Europa 74 20090 Cusago Milano Tel. +39 02.90394575 Fax +39 02.90394441 export@sirsafety.com www.sirsafety.com





Name: Gas alert BW CLIP CO

Code: FG110004

FG11 ELECTRICAL DEVICES GAS DETECTORS

## **Description**

The gas detector BW Clip CO is ideal for those who perform activities in hazardous environments. Operates for up to two years without maintenance: just turn on the device and it runs continuously - no need for calibration, sensor replacement, battery replacement or battery charging. With BW Clip it is possible to put the device in hibernation mode, by placing it in an appropriate housing whennot in use for more than a week, thus it extends its life by that period of time.

## It features:

- Maintenance-free: no replacement of sensors or batteries required.
- Compact, lightweight design with one-button operation.
- Designed to withstand harsh environments and extreme temperatures.
- Automated self-test of battery, sensor and electronics.
- Wide-angle flash, which alerts simultaneously with audible and vibrating alarm.
- Automatic logging of the 35 most recent gas events and bump test results.
- Water resistant
- Compatible with MicroDock II.
- Introduction of the automatic IntelliDox system which allows faster Bump Tests, reduced gas consumption, optional customisation based on operational needs.

Dimensions: 4,1 x 5,0 x 8,7 cm

Lightweight: 92 g

Humidity: 5%- 95% UR (non-condensing)

Visual, vibrating and acoustic alarm (95 dB); low/high.

Test: Activated detectors automatically performone internal diagnostic test

every 24 hours.

Ingress Protection: IP 66/67

Certifications : Class I, Div. 1, Group A, B, C, D Class I, Zone 0, Group IIC; II 1G; Ex ia IIC T4

DEMKO 14 ATEX 1356; IECEX: Ex ia IIC T4 IECEX UL 14.0063



Category	Colours	Available Sizes
		Made to order: U In stock:
Quantity/Carton	EC Standards	Important Information
	ATEX  Values:II 1G Ex ia IIC T4 Ga	Suitable to detect the presence of Carbon Monoxide.
1	IECEX  Values:Ex ia IIC T4 Ga	Make sure the product is in good condition, undamaged and clean before use. Once you become aware of a damage, you should immediately replace it.
		Keep in a cool dry place, far from heat sources.
Dexterity		Storage and cleaning