



SAFETY SYSTEM®
YOUR PROTECTOR

Sir Safety System S.p.A. Single-member company
Wholly-owned subsidiary of Sir Holding Srl
Via dei Fornaciaci, 9 - Santa Maria degli Angeli
06081 Assisi - Perugia - Italy
Tel.+39 075.8043737 - Fax +39 075.8048686

VAT Nr. IT 03359340548
Viale Europa 74
20090 Cusago Milano
Tel. +39 02.90394575
Fax +39 02.90394441

TECHNICAL SHEET

export@sirsafety.com
www.sirsafety.com



UNI EN ISO 9001:2015
UNI EN ISO 14001:2015

Name: GAS ALERT MAX XT II 4 GAS LEL/O2/CO/H2S	Code: FG110032	FG11 ELECTRICAL DEVICES - BW GAS DETECTORS
--	-----------------------	---

Description

Gas Alert Max XT II 4 Gas detects up to 4 gases simultaneously (Fuels, Oxygen, Sulphuric Acid, Carbon Monoxide).

- It constantly shows the gas concentration on the display.
- Full function self-test of sensor, battery status, circuit integrity and audio/visual alarms.
- Integrated sampling pump with a 3 m hose. for gas detection in confined spaces.
- Impact and water resistant casing.
- Advanced microprocessors, and easily plug in/replaceable catalytic and electrochemical sensors.
- Low, high, TWA (Time Weighted Average)
- STEL (Short Term Exposure Limit) alarm for toxic gases and OL (Over Limit) alarm for combustible gases, low battery.
- Acoustic (95dB), visual and vibrating alarm.
- Rechargeable batteries.
- Compact 13.1 x 7.0 x 5.2 cm.
- Weight 328 g.



ATEX CERTIFICATION

Category	Colours	Available Sizes
		U
Quantity/Carton	EC Standards	Important Information
1	EN 60079-0	Suitable to detect the presence of Fuels, Oxygen, Sulphuric Acid, Carbon Monoxide.
	EN 60079-11	Make sure the product is in good condition, undamaged and clean before use. Once you become aware of a damage, you should immediately replace it.
	EN 60079-26	Keep in a cool dry place, far from heat sources.
Dexterity		Storage and cleaning