



Sir Safety System S.p.A. Single-member company Wholly-owned subsidiary of Sir Holding Srl Via dei Fornaciai, 9 - Santa Maria degli Angeli 06081 Assisi - Perugia - Italy Tel.+39 075.8043737 - Fax +39 075.8048686 VAT Nr. IT 03359340548 Viale Europa 74 20090 Cusago Milano Tel. +39 02.90394575 Fax +39 02.90394441 export@sirsafety.com www.sirsafety.com





## Description

Gas Alert Max XT II 4 Gas detects up to 4 gases simultaneously (Fuels, Oxygen, Sulphuric Acid, Carbon Monoxide).

- It constantly shows the gas concentration on the display.
- Full function self-test of sensor, battery status, circuit integrity and audio/visual alarms.
- Integrated sampling pump with a 3 m hose. for gas detection in confined spaces.
- Impact and water resistant casing.
- Advanced microprocessors, and easily plug in/replaceable catalytic and electrochemical sensors.
- Low, high, TWA (Time Weighted Average)
- STEL (Short Term Exposure Limit) alarm for toxic gases and OL (Over Limit) alarm for combustible gases, low battery.
- Acoustic (95dB), visual and vibrating alarm.
- Rechargeable batteries.
- Compact 13.1 x 7.0 x 5.2 cm.
- Weight 328 g.

## ATEX CERTIFICATION

ATEX CERTIFICAT	ION	
Category	Colours	Available Sizes
		Made to order: U In stock:
Quantity/Carton  1	EC Standards	Important Information
	EN 60079-0	Suitable to detect the presence of Fuels, Oxygen, Sulphuric Acid, Carbon Monoxide.
	EN 60079-11	Make sure the product is in good condition, undamaged and clean before use. Once you become aware of a damage, you should
	EN 60079-26	immediately replace it.
		Keep in a cool dry place, far from heat sources.
	ATEX	
	Values:II 1G Ex da ia IIC T4 Ga	
	IECEX	
	Values:Ex ia IIC T4 Ga	
Doutouit		Character and algebraich
Dexterity		Storage and cleaning