

## **TECHNICAL SHEET**

**Sir Safety System S.p.A. Single-member company** Wholly-owned subsidiary of Sir Holding Srl Via dei Fornaciai, 9 - Santa Maria degli Angeli 06081 Assisi - Perugia - Italy Tel.+39 075.8043737 - Fax +39 075.8048686 VAT Nr. IT 03359340548 Viale Europa 74 20090 Cusago Milano Tel. +39 02.90394575 Fax +39 02.90394441

export@sirsafety.com www.sirsafety.com





FG11 ELECTRICAL DEVICES GAS Name: Gas alert MICRO 5 LEL/O2/CO/H2S/SO2 Code: FG110033 DETECTORS Description Gas Alert Micro 5 monitors up to 5 potential gases (Sulphuric Acid, Carbon Monoxide, Oxygen, Fuels, Sulphur Dioxide). ATEX CERTIFICATION 0.0 Category Colours Available Sizes Made to order: U In stock: Quantity/Carton EC Standards Important Information EN 60079-0 Suitable to detect the presence of Sulphuric Acid, Carbon Monoxide, Oxygen, Fuels, Sulphur Dioxide. EN 60079-1 Make sure the product is in good condition, undamaged and clean before use. Once you become aware of a damage, you should EN 60079-11 immediately replace it. Keep in acool dry place, far from heat sources. EN 60079-26 1 ATEX Values: II 1G Ex da ia IIC T4 or FHÍ ÈH»C Ga (M5), II 2 G Ex d ia IIC T4 or FH ÈH»C Gb (M5IR) IECEX Values:Ex ia IIC T4 Ga (M5), Ex d ia IIC Gb (M5IR) Dexterity Storage and cleaning